

ITEM	PART NUMBER	PART NAME	QTY
1	PD-01-010	LOT LABEL STOCK	1
2	PD-01-035	LOT THERMAL TRANSFER RIBBON	10 in

REVISIONS	
REV	DESCRIPTION
D	R&D Release
E	Add FCC number and MR Unsafe, Remove SGS
F	Add FCC statement

## Glean Urodynamics Male Delivery System

**CONTENTS**

1 Sensor  
1 Male Sheath  
1 Male Advancing Rod  2024-08-14

**REF** GUS-1000-M **LOT** 22222222

**S/N** 00000637



(01) 0 0860009 800649 (17) 240814 (10) 222222222 (21) 00000637









Rx Only




Not made with natural rubber latex.






FCC ID:2BHMUGUS1000

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference and (2) this device must accept any interference received including interference that may cause undesired operation.

**Bright Uro, Inc.**  
3 Goddard  
Irvine, CA 92618 USA  
TEL: +1 (949) 216-0873

**Patents:**  
[www.brighturo.com/patents](http://www.brighturo.com/patents)

Refer to Owner's Manual for Symbols Glossary. 

[www.brighturo.com/owners\\_manual](http://www.brighturo.com/owners_manual)

PD-01-005 Rev F

① ENTERED AT PRINTING STEP  
ASSIGNED BASED ON STERILITY DATE +6 MONTHS

② ENTERED AT PRINTING STEP

ENTERED AT PRINTING STEP

ENTERED AT PRINTING STEP



PART NAME:  
**POUCH LABEL, MALE**

UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES  
TOLERANCES: ANGLE  $\pm 1^\circ$  FRACTIONAL  $\pm 1/16$

INCHES X.X $\pm 0.1$	MILLIMETERS X. $\pm 2.5$
X.XX $\pm 0.01$	X.X $\pm 0.25$
X.XXX $\pm 0.005$	X.XX $\pm 0.13$
X.XXXX $\pm 0.0005$	X.XXX $\pm 0.013$

SIZE	PART NUMBER	REV
<b>A</b>	<b>PD-01-005</b>	<b>F</b>
DO NOT SCALE DRAWING		SHEET 1 OF 1