



Wipro GE Healthcare Pvt. Ltd.

John F Welch Technology Centre
Odyssey Building, 122, EPIP, Phase II
Whitefield, Bangalore 560 066
India
T 91 80 4088 1000
F 91 80 2841 1645
www.wiproge.com

Date: 18-Nov-2019

Federal Communications Commission
7435 Oakland Mills Road
Columbia, MD 21046

Subject: FCC Class II Permissive change application for Vscan Extend, Module FCC ID: **YOM-VSCANEXTEND** originally granted on Dec 04, 2016.

Vscan Extend Medical Equipment houses an off-the-shelf LAVA smartphone based on a customized Android OS. The processor and OS used in the smartphone is upgraded due to obsolescence. Smartphone's built in Wi-Fi feature is used for wireless communication.

The changes filed under this application are:

| SL.No | Parameter | Existing Equipment | New Equipment |
|-------|------------------|--------------------|--|
| 1 | Main Processor | MediaTek MT6735 | MediaTek MT6739WW |
| 2 | Wireless Module | MT6625L | MT6625L - Same as existing equipment |
| 3 | Operating System | Android 5.1 | Android 9.0 |
| 4 | RF Trace | - | Modified - length increased by 7.96mm due to new PCB layout. |
| 5 | Antenna Type | FPC - Dipole | FPC - Monopole |
| 6 | Gain | 1.2dBi | 1.2dBi - Same as existing equipment |

Sincerely,


(SREEDHAR J.V.)

(Signature)



Thank you for partnering with us on our mission of being at work for a healthier India

Wipro GE Healthcare Pvt. Ltd. Regd. Off : No. 4, Kadugodi Industrial Area, Bangalore 560 067. Karnataka India
CIN : U33111KA1990PTC016063 T 91 80 4180 1000 F 91 80 4180 1290